

UNIVERSITY OF MIAMI

ROSENSTIEL
SCHOOL of MARINE &
ATMOSPHERIC SCIENCE



Tritium Laboratory

3 June 2024

Tritium Laboratory
4600 Rickenbacker Causeway
Miami, Florida 33149-1031

Ph: 305-421-4100
Fax: 305-421-4112
E-mail: Tritium@rsmas.miami.edu

SWAB REPORT # 1094

SWAB DATE: 23 May 2024

Isotope Van # 625.3.08

Dr. James D. Happell
Associate Research Professor

Distribution:
SWAB Committee
Joe Lachmann
Tim Deering

COMMENTS TO SWAB REPORTS

15 December 2021

The LSC is now a Quantulus GCT 6220, with the SWAB counting assay having background cpm of 0.3 & 1.2 for ^3H & ^{14}C . This replaces an LSC with background cpm of 1.6 & 5.5 for ^3H & ^{14}C .

All samples are counted for 60 minutes, the instrument background is subtracted, and activities are reported in dpm/m^2 . Bucket blank activities are not subtracted. Counting errors (2 standard deviations) are also reported in dpm/m^2 . An error larger than the activity indicates that the activity is not significantly different from zero. All activities significantly above background will be in **bold**.

Criteria for SWAB Results

Category	^3H (dpm/m^2)	^{14}C (dpm m^2)	Recommendations
A	<500	<50	No action
B*	500-10,000	50-10,000	Needs cleaning before any natural tracer work. Decks in radiation vans with activities above $1000 \text{ dpm}/\text{m}^2$ should be cleaned.
C**	10,000-100,000	10,000-50,000	Must be cleaned before any use.
D***	>100,000	>50,000	May be a health hazard. Notify local radiation safety official.

Note: ^{14}C and ^{35}S have peak energies of 156 and 167 KeV, respectively; thus ^{35}S will be registered as ^{14}C by our counting techniques. Categories A, B and C are not a health hazard.

Recommended Cleaning Procedure

Wearing ordinary household rubber gloves:

^3H : Wash and scrub with radioactive cleanup detergent such as COUNT-OFF (50 ml COUNT-OFF to 4 liters of water), using sponges to distribute solution and reabsorb it.

^{14}C : Wash with 1% sulfuric or 2% hydrochloric (muriatic) acid with good ventilation (will dissolve carbonates, releasing $^{14}\text{CO}_2$). Follow up with wash as if for ^3H .

Disposal of Cleaning Materials (gloves, sponges, etc)

Categories A & B dispose as ordinary garbage, C & D contact your institution's radiation safety office.

Note: If category C or D is encountered, we try to notify the institution promptly by phone or email.

REPORT FOR SWAB # 1094

LOCATION: Lewes, DE

DATE: 23 May 2024

VESSEL: *Isotope Van # 625.3.08*

TECHNICIAN: Joe Lachmann

Sample #	Sample Identification	³ H dpm/m ²		¹⁴ C dpm/m ²	
		activity	error	activity	error
1	1st Vial Bkgnd	0	± 0	0	± 0
2	Initial bucket blank	11	± 68	-17	± 18
3	Tacky mat and floor near front door	-442	± 27	*5062	± 94
4	Middle floor	-361	± 23	*4686	± 90
5	Floor next to double doors	-206	± 17	*2812	± 71
6	Top of LSC	34	± 8	*310	± 27
7	Inside fume hood Stainless steel tray	25	± 20	15	± 13
8	Small sink and counter stainless steel tray	289	± 46	*122	± 18
9	Stainless steel tray above freezer/refrigerator	-536	± 28	*7022	± 110
10	Inside freezer	14	± 24	-2	± 2
11	Inside refrigerator	-6	± 181	14	± 15
12	Final bucket blank	3	± 6	-19	± 20

Comments

Please note that the error reported for each isotope is the two-standard deviation counting error. Reports may now contain values less than zero. Decay counting background samples will be distributed about the background vial, which means that negative values are possible. In the past we rounded the negative values to zero. Values are only significantly above background when they are positive and larger than the error. Please note that we are now using a Quantulus 6220 LSC which counts very near natural background. While the cleanup standards have not changed all values above background will now be in bold. This Isotope van was last used on the R/V Walton Smith in February and March of 2024 (see SWAB #1082). This van and the Walton Smith were contaminated with either ¹⁴C or ³⁵S during this cruise.

Figure 1
SWAB # 1094
23 May 2024

UNOLS Rad Van 625.3.08

